Oxygen Tank Duration Chart

Tank Cylinder Size:_____

Oxygen Liter Flow:_____

*******These are approximate durations*******

Fill or change your tank when regulator is in the Red Zone

Tank Size	Continuous Liter Flow *Decimal*	*Fraction*	Duration in Hours (Full Tank 2000PSI Green zone)
ML6 Tank *Homefill* *Approx. fill time when empty 75 MIN*	0.12	1/8	22 HRS 45 MIN
	0.25	1/4	10 HRS 45 MIN
	0.5	1/2	5 HRS 15 MIN
	0.75	3/4	3 HRS 30 MIN
	1		2 HRS 45 MIN
	2		1 HRS 15 MIN
	3		45 MIN
	4		30 MIN
	5		30 MIN
	6		15 MIN
ML9 Tank *Homefill* *Approx. fill time when empty 125 MIN*	0.12	1/8	34 HRS 15 MIN
	0.25	1/4	16 HRS 30 MIN
	0.5	1/2	8 HRS 15 MIN
	0.75	3/4	5 HRS 30 MIN
	1		4 HRS
	2		2 HRS
	3		1 HRS 15 MIN
	4		1 HRS
	5		45 MIN
	6		30 MIN
M6 Tank	0.12	1/8	22 HRS 45 MIN
	0.25	1/4	10 HRS 45 MIN
	0.5	1/2	5 HRS 15 MIN
	0.75	3/4	3 HRS 30 MIN
	1		2 HRS 30 MIN
	2		1 HR 15 MIN
	3		45 MIN
	4		30 MIN
	5		30 MIN
	6		15 MIN
D Tank ***Homefill and non Homefill*** *Approx filltime on homefill 3.5 HRS*	0.12	1/8	57 HRS 30 MIN
	0.25	1/4	27 HRS 45 MIN
	0.5	1/2	13 HRS 45 MIN
	0.75	3/4	9 HRS 15 MIN
	1		6 HRS 45 MIN
	2		3 HRS 15 MIN
	3		2 HRS 15 MIN
	4		1 HR 30 MIN
	5		1 HR 15 MIN
	6		1 HR
E Tank	0.12	1/8	3 DAYS
	0.25	1/4	1 DAY 10 HRS
	0.5	1/2	16 HRS
	0.75	3/4	11 HRS
	1		11 HRS 15 MIN
	2		5 HRS 30 MIN
	3		3 HRS 45 MIN
	4		2 HRS 45 MIN
	5		2 HR 15 MIN
	6		1 HR 45 MIN